



Attention!

Due to improved blocking composition in the microplates, the plates are more sensitive to appropriate washing.

Therefore, instead of washing the wells three times as stated in the assay procedure of previous Instructions (page 7), the wells should be washed **five times** with the provided wash solution.